

1mg~500mg	1	2	1	1	2	1	1	2	1										
1mg~50g	1	2	1	1	2	1	1	2	1	1	2	1	1	2	1				
1mg~100g	1	2	1	1	2	1	1	2	1	1	2	1	1	2	1				1
1mg~200g	1	2	1	1	2	1	1	2	1	1	2	1	1	2	1				1
1mg~500g	1	2	1	1	2	1	1	2	1	1	2	1	1	2	1				1
1mg~1kg	1	2	1	1	2	1	1	2	1	1	2	1	1	2	1				1
1mg~2kg(a)	1	2	1		2	1	1	2	1	1	2	1	1	2	1				1
1mg~2kg(b)	1	2	1	1	2	1	1	2	1	1	2	1	1	2	1				1
1mg~5kg	1	2	1	1	2	1	1	2	1	1	2	1	1	2	1				1
1mg~10kg	1	2	1	1	2	1	1	2	1	1	2	1	1	2	1				1
1mg~20kg(a)	1	2	1	1	2	1	1	2	1	1	2	1	1	2	1				1
1mg~20kg(b)	1	2	1	1	2	1	1	2	1	1	2	1	1	2	1				1

表1砝码最大允许误差的绝对值(|mpel|,以mg为单位)

标称值	e1	e2	f1	f2	m1	m2	m3
50kg	25	80	250	800	2500	8000	25000
20kg	10	30	100	300	1000	3000	10000
10kg	5	16	50	160	500	1600	5000
5kg	2.5	8	25	80	250	800	2500

2kg	1	3	10	30	100	300	1000
1kg	0.5	1.6	5	16	50	160	500
500g	0.25	0.8	2.5	8	25	80	250
200g	0.1	0.3	1	3	10	30	100
100g	0.05	0.16	0.5	1.6	5	16	50
50g	0.03	0.1	0.3	1	3	10	30
20g	0.025	0.08	0.25	0.8	2.5	8	25
10g	0.02	0.06	0.2	0.6	2	6	20
5g	0.016	0.05	0.16	0.5	1.6	5	16
2g	0.012	0.04	0.12	0.4	1.2	4	12
1g	0.01	0.03	0.1	0.3	1	3	10
500mg	0.008	0.025	0.08	0.25	0.8	2.5	
200mg	0.006	0.02	0.06	0.2	0.6	2	
100mg	0.005	0.016	0.05	0.16	0.5	1.6	
50mg	0.004	0.012	0.04	0.12	0.4		
20mg	0.003	0.01	0.03	0.1	0.3		
10mg	0.003	0.008	0.025	0.08	0.25		
5mg	0.003	0.006	0.02	0.06	0.2		
2mg	0.003	0.006	0.02	0.06	0.2		
1mg	0.003	0.006	0.02	0.06	0.2		